



## Blood and Transplant

### NHSBT Reagents

#### CERTIFICATE OF CONFORMANCE

**PRODUCT:** Weak anti-D control

**LOT NUMBER(S):** N046 3079

**EXPIRY:** 2020.03.26

**MANUFACTURE:** This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.

**PERFORMANCE:** The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.

The strength of this batch of weak anti-D control falls within the range  $0.09\text{iu} \pm 0.01$  as determined by a quantification analyser with traceable calibration to the WHO International Standard Anti-D immunoglobulin preparation available on the market at the time of testing.

**CERTIFICATION:** It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP.

**AUTHORISED BY:**

D A Wiles

**DATE:**

02/05/18

On behalf of Reagents Management Team

NHSBT Reagents is registered with MHRA to manufacture *In Vitro* Diagnostic Medical Devices and is certified by UL as the notified body

**NHSBT Reagents  
14 Estuary Banks  
Liverpool L24 8 RB**

**Tel: 0151 268 7157 Fax: 0151 268 7156**

**Email: [reagents@nhsbt.nhs.uk](mailto:reagents@nhsbt.nhs.uk) Web: [www.blood.co.uk/reagents](http://www.blood.co.uk/reagents)**