



Blood and Transplant
NHSBT Reagents

CERTIFICATE OF CONFORMANCE

PRODUCT: Papain

LOT NUMBER(S): N090 3328

EXPIRY: 2019.08.16

MANUFACTURE: This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.

PERFORMANCE: The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.

Cells papainised in parallel with this batch of NHSBT papain and NIBSC papain ref 92/658 have been tested against NIBSC anti-D ref 92/562 and found to have equivalent activity.

CERTIFICATION: It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP.

AUTHORISED BY:

DA Wilcox

DATE:

21/08/18

On behalf of Reagents Management Team

NHSBT Reagents is registered with MHRA to manufacture *In Vitro* Diagnostic Medical Devices and is certified by UL as the notified body

NHSBT Reagents
14 Estuary Banks
Liverpool L24 8 RB

Tel: 0151 268 7157 Fax: 0151 268 7156

Email: reagents@nhsbt.nhs.uk Web: www.blood.co.uk/reagents