

Blood and Transplant

NHSBT Reagents

CERTIFICATE OF CONFORMANCE

PRODUCT:

Weak anti-D control

LOT NUMBER(S):

N046 3106

EXPIRY:

2024.08.22

MANUFACTURE:

This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.

PERFORMANCE:

The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.

The strength of this batch of weak anti-D control falls within the range 0.09 ±0.01IU/mL as determined by a quantification analyser with traceable calibration to the WHO International Standard Anti-D immunoglobulin preparation available on the

market at the time of testing.

CERTIFICATION:

It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with

GMP.

AUTHORISED BY:

N-1000

DATE:

20/09/22

On behalf of Reagents Management Team

NHSBT Reagents is ISO13485 certified and is registered to manufacture CE marked In Vitro Diagnostic Devices

> **NHSBT Reagents** 14 Estuary Banks Liverpool L24 8RB

Tel: 0151 268 7157 Fax: 0151 268 7156

Email: reagents@nhsbt.nhs.uk Web: www.blood.co.uk/reagents