

## CERTIFICATE OF CONFORMANCE

PRODUCT:

**Papain** 

LOT NUMBER(S):

N090 3482

EXPIRY:

2025.05.12

MANUFACTURE:

This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood

Transfusion Services in the United Kingdom'.

PERFORMANCE:

The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and

performance features.

The activity of this batch of papain has been tested against reference material

and found to have equivalent activity to NIBSC papain ref 92/658

**CERTIFICATION:** 

It is certified that the above information is authentic and correct, the batch

DATE:

processing, packaging and analysis records were reviewed and found to be in

compliance with GMP.

**AUTHORISED BY:** 

2

14/06/24

On behalf of Reagents Management Team

NHSBT Reagents is ISO13485 certified and is registered to manufacture CE marked *In Vitro*Diagnostic Devices

NHSBT Reagents 14 Estuary Banks Liverpool L24 8RB

Tel: 0151 268 7157 Fax: 0151 268 7156

Email: reagents@nhsbt.nhs.uk Web: www.blood.co.uk/reagents