

INF1618/3 – Instructions for Use – Fluorescently labelled antibodies for FMH investigations



Blood and Transplant
Copy No:
Effective date: 14/06/2024



NHSBT Reagents,
14 Estuary Banks, Speke, Liverpool, L24 8RB,
Great Britain
Telephone: +44 (0)151 268 7157
Email: reagents@nhsbt.nhs.uk
www.blood.co.uk/reagents

Product Code	Product Name
9442FI	AEVZ 5.3 FITC Conjugated Control Reagent
9433FI	BRAD 3 FITC Conjugated Anti-RhD Reagent
9453PE	BIRMA 17C PE Conjugated Control Reagent

Amendments from the previous version of this IFU are in purple text.

Intended use

For professional use as an IVD device to give quantitative data to determine the size of a Feto-Maternal Haemorrhage (FMH).

BRAD 3 FITC is intended to be used to quantitate accurately the number of RhD positive cells in a mixture of RhD positive and negative cells and thereby estimate the size of an FMH by analysis of the maternal blood sample.

AEVZ 5.3 FITC is intended to be used as a negative control to quantitate accurately the number of RhD positive cells in a mixture of RhD positive and negative cells and thereby estimate the size of a Feto-Maternal Haemorrhage by analysis of the maternal blood sample.

BIRMA 17C PE is intended to be used to remove granulocytes during flow cytometry which may otherwise interfere with test results.

Principles of the examination method

Haemolytic Disease of the Fetus and Newborn (HDFN) can occur due to anti-D. Where an RhD negative pregnant person is carrying an RhD positive fetus, exposure to the D antigen via a Feto-maternal Haemorrhage (FMH) can stimulate anti-D production. HDFN due to anti-D can be prevented by administration of prophylactic anti-D post-partum or antenatally. The dosage of prophylactic anti-D required depends on the size of the FMH.

Red cells from an RhD negative sample with a suspected fetal bleed are incubated with a fluorochrome conjugated IgG monoclonal anti-D reagent, BRAD 3 FITC. This will bind to any RhD positive red cells present. Following incubation, the sample is tested by flow cytometry. Flow cytometry works on the principle of light scattering as cells pass through a laser beam in single file. The RhD positive cells, bound by BRAD 3 FITC can be counted due to their emission of fluorescence.

By flow cytometry, BRAD 3 FITC can be used to quantitate accurately the number of RhD positive cells in a mixture of RhD positive and negative cells, and thereby estimate the size of an FMH. The variability of fluorescence of rr cells (RhD negative) when labelled with BRAD 3 FITC means that a negative control antibody, such as AEVZ 5.3 FITC, should be used to gain accurate results. AEVZ 5.3 FITC is unreactive with human red cells irrespective of their RhD type.

Most FMH samples have elevated numbers of granulocytes, especially neutrophils. These granulocytes stain with BRAD 3 FITC non-specifically, giving FITC fluorescence that may overlap RhD negative and positive red cells. Labelling with BIRMA 17C PE enables the PE-labelled granulocytes to be eliminated from analysis by excluding cells with PE fluorescence.

INF1618/3 – Instructions for Use – Fluorescently labelled antibodies for FMH investigations



Blood and Transplant

Copy No:

Effective date: 14/06/2024

Components

- AEVZ 5.3 FITC is a human IgG3 engineered monoclonal antibody. Each vial contains 0.51mL +/- 0.05mL AEVZ 5.3 FITC conjugate in an amber bottle. The reagent has been prepared using a diluent containing 1% BSA and 0.099% sodium azide.
- BRAD 3 FITC is a human IgG3 anti-RhD monoclonal antibody. Each vial contains 0.51mL +/- 0.05mL BRAD 3 FITC conjugate in an amber bottle. The reagent has been prepared using a diluent containing 1% BSA and 0.099% sodium azide.
- BIRMA 17C is a mouse IgG1 anti-granulocyte antibody. Each vial contains 0.51 mL +/- 0.05mL BIRMA 17C R-Phycoerythrin (PE) conjugated reagent in an amber bottle. The reagent has been prepared using a diluent containing 1% BSA and 0.099% sodium azide.

Special materials and equipment required but not supplied.

- Calibrated volumetric pipettes.
- Flow cytometer
- Centrifuge
- Phosphate Buffered Saline Solution (PBS)
- Water bath or dry heat incubators
- Test tubes which are demonstrated to be non-adherent for red cells, e.g., polyethylene.

Reagent Preparation

Allow to reach room temperature, mix before use.

Storage and shelf life after first opening

Store at 2-8°C in the dark.

Once opened the device can be used until stated expiry date.

Keep stored in the amber bottle.

Do not use beyond the expiry date.

After use, the vial should be capped and returned to the correct storage conditions and temperature.

Warnings and precautions

The device and any contaminated packaging should be disposed in accordance with local state or national legislation.

For healthcare professional use only.

The reagent should not be used if turbid or if a precipitate, gel or particles are present.

This device is not provided sterile.

Do not use if the reagent vial is cracked or leaking.

It is imperative to use accurate, properly calibrated volumetric pipettes to avoid variations which may affect the test outcome.

When used in accordance with the Instructions for Use and Good Laboratory Practices, there is limited potential for carryover.

The fluorescent activity of FITC and PE conjugated reagents is degraded by exposure to light. Minimise the exposure of these reagents to light where possible.

Where a diluted working solution is made up from the original vial (see examination procedure methods A and B) use immediately. This must be discarded once the testing is completed. Do not store diluted working solutions for further use.

It is essential that samples are thoroughly mixed before the required cell aliquot is removed for testing.

Minimum washing of test samples should be applied to prevent the loss of fetal cells.

Do not freeze.

Primary sample collection, handling, and storage

Use EDTA samples according to the current edition of the [British Society for Haematology \(BSH\)](#) Guidelines for the Estimation of Fetomaternal Haemorrhage.

Controlled if copy number stated on document and issued by QA

(Template Version 03/02/2020)

Examination procedure

The following considerations should be derived from, and performed in accordance, with the current BSH guidelines for the estimation of Fetomaternal Haemorrhage:

- Sample preparation (mixing, washing, centrifugation)
- Aliquot testing
- Controls
- Number of events collected and analysed.
- Calculation of FMH

Method A: BRAD 3 FITC and AEVZ 5.3 FITC (1:10 working solution)

The recommended red cell concentration for all samples and controls is 5%.

1. Example preparation of a 1:10 working dilutions for use of BRAD 3 FITC and AEVZ 5.3 FITC
 - Add 40µL of BRAD 3 FITC to 360µL PBS.
 - Add 40µL of AEVZ 5.3 FITC to 360µL PBS.
 - Mix thoroughly (e.g. Vortex mixer for 2 seconds)
 - Solutions are ready to use.
2. Test samples and controls (BRAD 3 FITC and AEVZ 5.3 FITC to be run in parallel)
 - Add 50µL of diluted BRAD 3 FITC *or* diluted AEVZ 5.3 FITC to 50µL of a well mixed 5% cell suspension to be tested.
 - Mix gently and cap tubes.
 - Incubate tests at 37°C in the dark for 30 min.
 - Wash the cells once with PBS, discard the supernatant carefully ensuring no cells are discarded.
 - Resuspend the cell button in 800-1000µL PBS.
 - Samples are ready for Flow Cytometer analysis.

Method B: BRAD 3 FITC and AEVZ 5.3 FITC with BIRMA 17C PE (1:10 working solution)

The recommended red cell concentration for all samples and controls is 5%.

1. Example preparation of a 1:10 working dilutions for use of BRAD 3 FITC and AEVZ 5.3 FITC **with** BIRMA 17C PE
 - Add 20µL of BRAD 3 FITC *and* 20µL of BIRMA 17C PE to 160µL PBS.
 - Add 20µL of AEVZ 5.3 FITC *and* 20µL of BIRMA 17C PE to 160µL PBS.
 - Mix thoroughly (e.g., Vortex mixer for 2 seconds)
 - Solutions are ready to use.
2. Test samples and controls (BRAD 3 FITC/ BIRMA 17C PE and AEVZ 5.3 FITC / BIRMA 17C PE to be run in parallel)
 - Add 50µL of diluted BRAD 3 FITC/ BIRMA 17C PE *or* diluted AEVZ 5.3 FITC/ BIRMA 17C PE to 20µL of a well mixed 5% cell suspension to be tested.
 - Mix gently and cap tubes.
 - Incubate tests at 37°C in the dark for 30 min.
 - Wash the cells once with PBS, discard the supernatant carefully ensuring no cells are discarded.
 - Resuspend the cell button in 800-1000µL PBS.
 - Samples are ready for Flow Cytometer analysis.

Method C: reagents used directly from the original vial.

1. Procedure for use of BRAD 3 FITC and AEVZ 5.3 FITC without BIRMA 17C PE

BRAD 3 FITC is supplied ready for use at 5µl per test.

- Sample cells should be resuspended in a final volume of 50µl PBS prior to testing. For example, dispense 20µl of 3-5% erythrocytes into a tube and add 30µl of PBS.
- Add 5µl of BRAD 3 FITC.
- Mix and incubate at 37°C in the dark for 30 min.
- Wash the cells once in PBS, discard the supernatant carefully ensuring no cells are discarded.
- Resuspend the cell button in 800-1000µL PBS.
- Samples are ready for Flow Cytometer analysis.

AEVZ 5.3 FITC is supplied ready for use at 5µl per test as a negative control alongside BRAD 3 FITC.

- Sample cells should be resuspended in a final volume of 50µl PBS prior to testing. For example, dispense 20µl of 3-5% erythrocytes into a tube and add 30µl of PBS.
- Add 5µl of AEVZ 5.3 FITC.
- Mix and incubate at 37°C in the dark for 30 min.
- Wash the cells once in PBS, discard the supernatant carefully ensuring no cells are discarded.
- Resuspend the cell button in 800-1000µL PBS.
- Samples are ready for Flow Cytometer analysis.

Method D: reagents used directly from the original vial.

1. Procedure for use of BRAD 3 FITC and AEVZ 5.3 FITC with BIRMA 17C PE

BIRMA 17C PE is ready for use at 5 µl per test.

- Sample cells should be resuspended in a final volume of 60 µl PBS prior to testing. For example, dispense 20 µl of 3-5% erythrocytes into a tube and add 40 µl of PBS. Add 5 µl of either BRAD 3 FITC or AEVZ 5.3 FITC and 5 µl of BIRMA 17C PE.
- Mix and incubate at 37°C in the dark for 30 min.
- Wash the cells once in PBS, discard the supernatant carefully ensuring no cells are discarded.
- Resuspend the cell button in 800-1000µL PBS.
- Samples are ready for Flow Cytometer analysis.

Control procedure

Users are responsible for determining the appropriate quality control procedures for their laboratory and for complying with applicable laboratory regulations. If controls set up with the batch of tests fail to give required results, then all tests must be repeated.

A negative control such as AEVZ 5.3 FITC reagent should be included to control for variability in fluorescence of rr (RhD negative) cells (methods A, B, C and D). PE conjugated BIRMA 17C can be used in conjunction with BRAD 3 FITC in a dual colour assay for FMH quantitation (methods B and D).

Interpretation of results

Result analysis and calculation of the FMH should be performed according to the current BSH Guidelines on the Estimation of Fetomaternal Haemorrhage. For NHSBT centres please refer to the Management Process Description MPD444. See Figure 1 for identification of R1r cells stained with BRAD 3 FITC and R1r and rr cells stained with AEVZ5.3 FITC. Figure 2 shows how granulocytes stained with BIRMA 17C PE can be excluded from analysis when the dual colour assay is employed.




Figure 1. AEVZ 5.3 FITC with R1r and rr cells (blue peak) and BRAD 3 FITC with R1r cells (red peak)




Figure 2. **A:** gate A excludes debris (with low FS and SS) and the cells within gate A are displayed in a SS vs. FL2 dot plot (**B**) where gate B excludes PE-positive cells stained with BIRMA 17C PE. Cells in gate B are displayed as a histogram (**D**) where a marker can be applied to cover the range of fluorescence of D-positive cells and their percentage is calculated by the flow cytometer; example shown 0.90%. **C:** dot plot of BRAD 3 FITC vs. BIRMA 17C PE shows D-positive cells (green/FITC) and granulocytes (purple/PE). A gate to exclude FL2-positive cells may alternatively be placed on a dot plot to generate the histogram, or a quadrant may be applied as shown in C and the percentage of cells in the lower right quadrant determined.

Performance characteristics

BRAD 3 FITC: Controls are prepared containing RhD positive cells at 100%, 1.0% and 0.20% of the total red cell population. When the fluorescently labelled antibodies for FMH investigation are used together to test these controls for RhD positive cells, results are given within the following limits.

Nominal % of RhD positive cells	Result range from FMH antibody investigation.
100%	>99.2%
1.0%	0.91%-1.15%
0.20%	0.16%-0.23%

BIRMA 17C PE: The antibody specificity is confirmed as anti-CD66b. The reagent is positive with granulocytes and negative with lymphocytes by flow cytometry.

Limitations of the examination procedure

BRAD 3 FITC is not for use as a blood grouping reagent.

BRAD 3 FITC binds all RhD positive red cells tested except those of the rare DVI or R₀^{Har} types.

FMH involving D phenotypes expressing fewer than 1000 RhD antigens per red cell, may not be detected.

Most FMH samples have elevated numbers of granulocytes, especially neutrophils. These granulocytes stain with **BRAD 3 FITC** and other IgG antibodies non-specifically, giving FITC fluorescence that may overlap RhD negative and RhD positive red cells. Labelling with **BIRMA 17C PE** enables the PE-labelled granulocytes to be eliminated from analysis by excluding cells with PE (FL-2) fluorescence (**methods B and D**).

False positive or false negative results may occur due to contamination of test material, improper storage, incorrect incubation time or temperature, improper or excessive centrifugation or deviation from the recommended technique.

Visual evidence of hyperlipidaemia or haemolysis and age of specimen may affect the interpretation of test results.

Prophylactic anti-D binds to RhD antigen. This action may give an incorrect (lower estimated FMH) result with BRAD 3 FITC. The timing of the last dose of prophylactic anti-D given before the FMH sample should be noted.

Literature references

Directive 98/79/EC on In vitro diagnostic medical devices.

BSH guidelines for the estimation of fetomaternal haemorrhage (FMH)

Lloyd-Evans et al, (1996), Use of a directly conjugated monoclonal anti-D (BRAD 3) for quantification of fetomaternal hemorrhage by flow cytometry, *Transfusion*, 36, 432-437.

Lloyd-Evans et al, (1999), Detection of weak D and DVI red cells in D-negative mixtures by flow cytometry: implications for feto-maternal haemorrhage quantification and D typing policies for newborns, *British J. Haematol.* 104 621-625.

Kumpel et al, (2014) Accurate quantitation of D+ feto-maternal haemorrhage by flow cytometry using a novel reagent to eliminate granulocytes from analysis. *Transfusion*, 54, 1305–1316.

Lloyd-Evans P, Austin EB, Gilmour JEM, Scott ML (1999) Use of a negative control antibody in the quantitation of feto-maternal haemorrhage by flow cytometry. *Transfusion Medicine* 9 suppl. 1:33. 9.

Kumpel et al, (2014) Accurate quantitation of D+ fetomaternal hemorrhage by flow cytometry using a novel reagent to eliminate granulocytes from analysis. *Transfusion*, 54, 1305–1316.

INF1618/3 – Instructions for Use – Fluorescently labelled antibodies for FMH investigations



Blood and Transplant

Copy No:

Effective date: 14/06/2024

Manufacture

These reagents are manufactured by IBGRL at NHSBT 500 North Bristol Park, Northway, Filton, Bristol BS34 7QH. Great Britain. Phone +44 (0)117 921 7500 on behalf of NHSBT Reagents.

Note – Any serious incident that has occurred in relation to using this reagent should be reported to the manufacturer and the competent authority in which the user and/or the patient is established.

EC	REP
----	-----

Quality First International OÜ, Laki 30, 12915 Tallinn, Estonia

INF1618/3 – Instructions for Use – Fluorescently labelled antibodies for FMH investigations



Blood and Transplant

Copy No:

Effective date: 14/06/2024

Symbols used on NHSBT Reagents labels.

Note - not all symbols listed are applicable for this product - please refer to product labels.

Detail	Label details
Batch code symbol	
Use by date symbol	
Expiry date format	YYYY.MM.DD
In Vitro Diagnostic medical device symbol	
Instructions for use symbol (with website - electronic IFU)	 blood.co.uk/reagents
Negative control symbol	
Positive control symbol	
EC Rep symbol	

Detail	Label details
2-8°C temperature range symbol	
Below -20°C symbol	
CE Mark symbol	
UKCA symbol	
Manufacturer's symbol	
Keep Away from Sunlight symbol	
Contains human blood or plasma derivatives symbol	
Unique Device Identifier symbol	

Lot number Format

Lot numbers are in the following format:

For BRAD 3 FITC (9433FI) – FBr3.XX

For AEVZ 5.3 FITC (9442FI) – FAEVZ5.3-XX

For BIRMA 17C PE (9453PE) – PBM17C.XX

XX Lot number