



Blood and Transplant
NHSBT Reagents

CERTIFICATE OF CONFORMANCE

PRODUCT: Anti-D 0.5 IU / mL

LOT NUMBER(S): N186 3019

EXPIRY: 2026.09.25

MANUFACTURE: This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.

PERFORMANCE: The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.

The strength batch of this batch of anti-D control has been determined by a quantification analyser with traceable calibration to the WHO International Standard Anti-D immunoglobulin preparation available on the market at the time of testing.

CERTIFICATION: It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP.

AUTHORISED BY:


On behalf of Reagents Management Team

DATE:

29/10/2024

NHSBT Reagents is ISO13485 certified and is registered to manufacture CE marked *In Vitro* Diagnostic Devices

NHSBT Reagents
14 Estuary Banks
Liverpool L24 8RB

Tel: 0151 268 7157 Fax: 0151 268 7156

Email: reagents@nhsbt.nhs.uk Web: www.blood.co.uk/reagents