

## **CERTIFICATE OF CONFORMANCE**

PRODUCT:	Papain
LOT NUMBER(S):	N090 3509
EXPIRY:	2026.06.01
MANUFACTURE:	This product complies with the requirements of Directive 98/79/EC and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.
PERFORMANCE:	The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.
	The activity of this batch of papain has been tested against reference material and found to have equivalent activity to NIBSC papain ref 92/658
CERTIFICATION:	It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP.
AUTHORISED BY:	N POQ DATE: 09/07/25
	On behalf of Reagents Management Team
NHSBT Reagents is ISO13485 certified and is registered to manufacture CE marked In Vitro Diagnostic Devices	
NHSBT Reagents 14 Estuary Banks Liverpool L24 8RB Tel: 0151 268 7157 Fax: 0151 268 7156 Email: <u>reagents@nhsbt.nhs.uk</u> Web: www.blood.co.uk/reagents	

Т