



CERTIFICATE OF CONFORMANCE

PRODUCT: Allo Absorption 2 cell set

LOT NUMBER(S): R405 3670

EXPIRY: 2026.01.29

MANUFACTURE: This product complies with the requirements of the UK MDR 2002 (as amended) and the recommendations contained in the current edition of 'Guidelines for the Blood Transfusion Services in the United Kingdom'.

PERFORMANCE: The Quality Control testing of this product confirms compliance with all applicable guidelines and regulations and that it meets all specified design and performance features.

CERTIFICATION: It is certified that the above information is authentic and correct, the batch processing, packaging and analysis records were reviewed and found to be in compliance with GMP.

AUTHORISED BY:


On behalf of Reagents Management Team

DATE:

19/12/2025

NHSBT Reagents is ISO13485 certified and is registered to manufacture UKCA marked *In Vitro* Diagnostic Medical Devices

NHSBT Reagents

14 Estuary Banks

Liverpool L24 8 RB

Tel: 0151 268 7157 Fax: 0151 268 7156

Email: reagents@nhsbt.nhs.uk Web: www.blood.co.uk/reagents