



**To: Transfusion Laboratory Managers**

**26 January 2026**

**Field Safety Notice: FSN 2025-05 FastQ B27 HLA Testing Update**

Dear Colleague,

Further to the communication issued on 4<sup>th</sup> December 2025 regarding the supplier field safety notice. We have completed retesting patients where an original blood sample was available and testing samples from rebleeds when received. **All** results are concordant with the original and the reports issued are correct.

We are pleased to communicate that we have completed a successful verification of the revised supplier instructions, and we will resume normal testing by the week commencing 9<sup>th</sup> February 2026. From this date, we will return to the report format that you are familiar with.

If you have any queries, please direct them to [HINATNL@nhsbt.nhs.uk](mailto:HINATNL@nhsbt.nhs.uk).

Thank you for your patience.

**Please forward this letter to laboratory and clinical staff as appropriate.**

**Dr Katy Latham PhD FRCPath**

*Assistant Director of Histocompatibility and Immunogenetics*

*NHS Blood and Transplant*

